

# Protek DT201/DT300/DT628A/JT6

Digital Multimeter

**Protek**  
Measurement and Control




Protek DT201

Function and Characteristic	Range	Accuracy
DC V	200mV ~ 500V	± (0,5% + 2)
AC V	200V ~ 500V	± (1,2% + 10)
DC A	2000µA ~ 200mA	± (1% + 2)
Resistance	200Ω ~ 2000KΩ	± (0,8% + 2)
Diode Test	Yes	
Battery Check	9V	
Power Supply	12V, A23	
Max. Display	1999	



Protek DT300

Function	Range	Accuracy
DC V	2V ~ 500V	± (0,5% + 2)
AC V	200V ~ 500V	± (1,2% + 10)
DC A	200mA	± (1,2% + 2)
Resistance	2KΩ ~ 2000KΩ	± (1,0% + 2)
Diode Test	Yes	
Continuity Buzzer	Yes	
Power Supply	12V, A23	
Max. Display	1999	



Protek DT628A

Function	Range	Accuracy
DC V	2000mV ~ 500V	± (0,5% + 2)
AC V	200V ~ 500V	± (1,2% + 10)
DC A	200mA	± (1,2% + 2)
Resistance	2000Ω ~ 2000KΩ	± (1,0% + 2)
Diode Test	Yes	
Power Supply	12V, A23	
Max. Display	1999	



Protek JT6

Function	Range	Accuracy
DC V	320mV ~ 500V	± (1,3% + 4)
AC V	3,2V ~ 500V	± (2,0% + 4)
DC A	320µA ~ 320mA	± (1,5% + 5)
AC A	320µA ~ 320mA	± (2,5% + 5)
Resistance	320Ω ~ 320KΩ	± (2,0% + 4)
	3,2MΩ	± (6,0% + 2)
	30MΩ	± (9,0% + 5)
Diode Test	Yes	
Continuity Buzzer	Yes	
Power Supply	Two LR-44	
Data Hold	Yes	
Auto Power Off	Yes	
Max. Display	3200	

Digital  
Multimeter

TEL:+82-32-870-5600 FAX:+82-32-870-5640 E-mail:sale@gsinstrument.com http://www.gsinstrument.com GSI

Messtechnik & Elektronik GmbH **mem**

Beratung und Verkauf  
elektronischer Messgeräte  
Pilartzstraße 9

Tel: 08071 / 92306-0  
www.mem-gmbh.de

FAX: 08071 / 92306-19  
mail@mem-gmbh.de